



Product Service

CERTIFICATE

No. Z1 13 08 39801 004

Holder of Certificate: Invacare Portugal Lda

Rua Estrada Velha, 949
4465-784 Leça do Balio
PORTUGAL

Certification Mark:



Product: Medical Beds

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713026196

Date, 2013-08-12
Page 1 of 2

(Torsten Zimmer)





Product Service

CERTIFICATE
No. Z1 13 08 39801 004

Model(s): **Alegio NG**

Parameters:

Voltage:	230 VAC
Frequency:	50 Hz
Current:	max. 1.0 A
Degree of protection:	IPX4 or IPX6
Protection class:	II
Safety working load:	170 kg
Max. patient weight:	135 kg
Operating time (INT):	10% (2 min/18min)
Applied part:	Type B
Application environment:	3, 4
Total width:	90 cm
Total length:	211 cm
Range of adjustment of the height of the mattress support platform:	28 - 80 cm
Dimensions of mattress support platform:	90 x 200 cm
Mattress support platform:	4-sectioned

Tested according to:

- EN 60601-1:2006
- EN 60601-1-11:2010
- EN 60601-1-2:2007
- EN 60601-2-52:2010
- EN 60601-1-6:2010
- EN 62366:2008

Production Facility(ies): 39801